Metro Scale Company, Inc

Certificate of Calibration

17670 Allen Rd, Melvindale
MI, 48122
Phone: 313.382.6190
Fax: 313.382.6194

PJLA Calibration Accreditation #59303

Customer: Daytona Ready Mix
Address: 7520 23 Mile Rd
City, State Zip: Shelby Twp, MICHIGAN 48316

Certificate ID: 3776681260104
ISO Number:
Date 4/1/2026

Indicator Mfr.	Base Mfr.	Cal Date	Scale ID	Scale Location
Rinstrum	N/A	4/1/2026	3776681	Portable Plt #2 Water Scale
Indicator Model	**Base Model**	**Due Date**	**Scale Class**	**Scale Range**
BL320	N/A	10/31/2026	III	0 - 4000 lb x 5 lb
Indicator Serial	**Base Serial**	**Procedure**	**Scale Status**	
3776681	N/A	HB-44	In Service	
Test Interval				
6 Months				

EQUIPMENT CONDITIONS

[Working] Non-Working Clean Dirty Out of Level Out of Service

All tolerances calculated in conformance with Handbook 44 Table 6.

LOAD TEST

Preliminary Load Test		
Test Wt.	**Reading**	**Error**
0 lb	0 lb	0 lb
1000 lb	1000 lb	0 lb
2000 lb	2000 lb	0 lb

In maintenance tolerance? **[Yes]** No N/A

Final Load Test		
Test Wt.	**Reading**	**Error**
0 lb	0 lb	0 lb
1000 lb	1000 lb	0 lb
2000 lb	2000 lb	0 lb

In maintenance tolerance? **[Yes]** No N/A

TEST INFORMATION

Test Weight Classification:	F
Traceability Certificate Number(s):	MI-03-24-16753 Cal Date: 3/29/2024 Recal Date: 3/29/2026
Standards Used:	101, 102
Expanded Uncertainty:	2300 grams
Test Location:	**[Onsite]** Offsite
Overall Result:	**[Pass]** Fail Adjust
Was the scale within customers required accuracy?	**[Y]** N N/A
Environmental Conditions:	**[Acceptable]** Unacceptable
Temperature:	38°F
Humidity:	50%
Comments / Notes:	Dubay Dr. Mt Clemens
Technician:	Michael Lulko

Certifying Technician:

Scales were calibrated with certified test weights. Adjustments made to restore and/or maintain the accuracy of the scale conform to the tolerances established by the National Institute of Standards and Technology as specified in Handbook 44 Section 2.2, or Manufacturers Specifications. Best measurement of uncertainty calculated using a coverage factor of K=2. This provides confidence level of 95%. This certificate shall not be reproduced, except in full, with the written approval of the laboratory.